

912-0010 NC-200 IQ/OQ Kit

Contents	P/N: 941-0019	Test Cassette 1 with one dot.
	P/N: 941-0020	Test Cassette 2 with two dots.
	P/N: 071-0022	1 bottle with 1.3 mL of fluorescent beads for NC-200 IQ/OQ Kit . The solution contains multicoloured 10 µm beads, Tween-20 (0.02%) and NaN ₃ (2mM ~ 0.01%) in H ₂ O. CAS no. for Tween-20 is 9005-64-5 CAS no. for NaN ₃ is 26628-22-8 EINECS no. for NaN ₃ is 247-852-1
Application	NC-200 IQ/OQ Kit is part of the <i>NucleoCounter</i> [®] <i>NC-200™</i> system and is used for performing an Installation Qualification (IQ) and an Operation Qualification (OQ).	
Principle	NC-200 IQ/OQ Kit contains multicoloured particles that allow testing of the performance of all the channels of the <i>NucleoCounter</i> [®] <i>NC-200™</i> .	
Usage	With respect to description of the detailed procedure, please refer to the appropriate application and technical notes, and user manual for the <i>NucleoCounter</i> [®] <i>NC-200™</i> instrument and the <i>NucleoView™ NC-200™ Software</i> . NC-200 IQ/OQ Kit is for research and development purposes only and is not for diagnostic or therapeutic use.	
Storage	The fluorescent bead solution should be stored at 2-8°C. Protect against light.	
Stability	Expiry dates are shown on the bead bottle and outside primary packaging. The NC-200 IQ/OQ Kit is produced 15 months before expiry date.	
Safety Information	Fluorescent bead solution: In case of contact with eyes, rinse immediately with plenty of water and seek medical advice. If a spill is observed, perform a clean-up of the area, which may have been in contact with the solution. Use gloves and suitable protective clothing. Please also refer to SDS regarding safety information.	
Disposal of Waste	After use, the NC-200 IQ/OQ Kit should be disposed of according to national or local laws and regulations regarding the nature of the mixture it contains.	