

912-0021 NC-3000 IQ/OQ Kit

Content	P/N: 071-0041 One vial with 1.3 mL fluorescent 10 µm beads. Buffer solution: 0.02% Tween-20 and 2 mM ~ 0.01% NaN ₃ in H ₂ O. CAS no. for Tween-20: 9005-64-5 CAS no. for NaN ₃ : 26628-22-8 EINECS no. for NaN ₃ : 247-852-1
Application	The NC-3000 IQ/OQ Kit is part of the NucleoCounter® NC-3000™ system and is used when conducting an Installation Qualification (IQ) and an Operation Qualification (OQ).
Principle	The NC-3000 IQ/OQ Kit contains multicoloured beads for testing the performance of all the channels of the NucleoCounter® NC-3000™.
Use	For a detailed IQ/OQ procedure, please refer to the appropriate technical note and user manuals for the NucleoCounter® NC-3000™ instrument and the NucleoView™ Software. The NC-3000 IQ/OQ Kit is for research and development use only and not for diagnostic or therapeutic uses.
Storage	The fluorescent bead solution should be store at 2-8°C. Protect against light.
Stability	The shelf life for the kit is 15 months from the production date. The expiry date is shown on the kit and on the bead vial labels.
Safety Information	Fluorescent bead solution: In case of eye contact, rinse immediately with plenty of water and seek medical advice. If a spill is observed clean the area, which may have been in contact with the solution according to national or local laws and regulations. Use gloves and suitable protective clothing. Refer to the SDS for full safety information.
Waste Disposal	After use, the NC-3000 IQ/OQ Kit should be disposed of according to national or local laws and regulations regarding the nature of the mixture it contains.